

T40 II

Portable Single Gas Monitor



Product Manual

*The Essential Guide for
Safety Teams and
Instrument Operators*

Edition: 1.0

Sep 12, 2022

Part Number: 17160575-1

INDUSTRIAL
SCIENTIFIC

Content

1. Certifications	2
2. Warnings and cautionary statements	3
3. Product Overview	3
3.1 Product Specifications.....	4
3.2 Hardware Overview.....	5
4. Operation and Instructions	6
4.1 Display Overview.....	6
4.2 Start-up, Shutdown and Always on.....	8
4.3 Configuration mode.....	8
4.4 Operation mode.....	13
4.5 Bump test (or "functional test").....	15
4.6 Zeroing.....	16
4.7 Calibration.....	17
5. Alarms and Notifications	18
6. Replace the battery	19
7. Warranty Policy	20
8. Limitation of Liability	20

1. Certifications

Directive or CB	Area Classifications	Approved Temperature Range	Standards
IECEX	Ex ia IIC T4 Ga Ex ia I Ma	-40 °C to +50 °C -40 °F to +122 °F	IEC 60079-0:2017 Ed.7.0 IEC 60079-11:2011 Ed.6.0
ATEX	Ex ia IIC T4 Ga Ex ia I Ma	-40 °C to +50 °C -40 °F to +122 °F	EN IEC 60079-0:2018 EN 60079-11:2012
UL	Class I, Division 1, Groups A, B, C, D T4 Class I, Zone 0, AEx ia IIC T4 Ga	-40 °C to +50 °C -40 °F to +122 °F	ANSI/UL 60079-0-2020 Ed.7 ANSI/UL 60079-11-2018 Ed.6 ANSI/UL 913-2019 Ed.8 UL 61010-1 3rd Edition (2012), AMD1: 2018
CSA	Ex ia IIC T4; Class I, Division 1, Groups A, B, C, D T4	-40 °C to +50 °C -40 °F to +122 °F	CAN/CSA C22.2 No. 60079-0:19 CAN/CSA-C22.2 No. 60079-11:14 (r2018) CAN/CSA C22.2 No.61010-1-12, UPD1: 2015, UPD2: 2016, AMD1: 2018
China Ex	Ex ia IIC T4 Ga	-40 °C to +50 °C -40 °F to +122 °F	GB/T 3836.1- 2021 GB/T 3836.4- 2021 GB/T 4208-2017
China CPA	Metrological	-10 °C to +40 °C 14 °F to +104 °F	JJF 1363-2019 JJF 1421-2013
China MA	Ex ia I Ma	-20 °C to +40 °C -4 °F to +104 °F	MT 703-2008 GB/T 3836.1-2021 GB/T 3836.4-2021

2. Warnings and cautionary statements



For maximum safety and optimal performance, read and understand the manual before operating or servicing the unit. Failure to perform certain procedures or note certain conditions may impair the performance of this product.



For safety reasons, this equipment must be operated and serviced by qualified personnel only. Customer site assembly is not recommended, improper disassembly may reduce the performance of the instrument.



Substitution of components may impair intrinsic safety and may cause an unsafe condition.



DO NOT REPLACE BATTERY WHEN AN EXPLOSIVE ATMOSPHERE IS PRESENT. USE ONLY RAMWAY ER14335 BATTERIES.



Before use, ensure that the ESD film on the LCD display and label is not damaged or peeling.



Obstruction of sensor openings—due to dust, dirt, water, or another cause—can inhibit the unit's ability to measure gas concentrations accurately. When this occurs, readings may appear lower than the actual gas concentration. Keep sensor openings clean, dry, and properly exposed to the ambient air.



Obstructed, contaminated, or damaged sensor water barriers (or their gaskets) can inhibit the unit's ability to measure gas concentrations accurately. When this occurs, readings may appear lower than the actual gas concentration. Contact after-sales service to replace the sensor water barriers and gaskets as needed.



Service the unit, use its communications port, and change its battery cell only in nonhazardous locations. Not for use in oxygen-enriched atmospheres.



Contact your service representative immediately if you suspect that the unit is working abnormally.

3. Product Overview

T40 II single gas monitor is compact, easy to carry and use. It is specially designed to detect the concentration toxic gases, and mainly used in Petrochemical industry and coal mine. The monitor can continuously monitor the concentration of harmful gas in the surrounding environment, once the gas concentration reaches the low and high alarm and TWA/STEL alarm preset values, T40 II will produce sound, light and vibration alarms

3.1 Product Specifications

Item	Description	
Display	Segment LCD	
Keypad buttons	Two buttons	
Technology	Electrochemical	
Battery Pack	Replaceable RAMWAY ER14335 3.6 V Primary Lithium-thionyl chloride (Li-SOCl ₂) battery cell, 1.65AH, 2/3AA. Two-years ^a run time	
Case materials	polycarbonate with a protective rubber over-mold	
Alarms	Visual alarm (two red LEDs). Audible alarm(95 dB). Vibration alarm	
Ranges	Carbon Monoxide (CO)	Hydrogen Sulfide(H ₂ S)
	0-1000 ppm	0-100 ppm
	0-2000 ppm	0-200 ppm
		0-500 ppm
Display resolution	1 ppm	0.1 ppm
Calibration gas and concentration	100 ppm CO	25 ppm H ₂ S
Response Time	<10s	<10s
Dimensions	82 x 60 x 27 mm (3.23 " x 2.36 " x 1.07 ")	
Weight	85 g (3.0 oz.), typical	
Ingress protection	IP66 / IP68	
Operating Temp ^c	-40 °C to +50 °C (-40°F to +122 °F)	
Operating humidity	15 to 95% relative humidity (RH) noncondensing (continuous)	
Pollution Degree	2	
Maximum use altitude	5000m above sea level	



^a Depending on operating conditions; the amount of time the unit is in alarm; and the enablement of unit's confidence indicator.

^b Apply when the instrument is calibrated using the stated calibration gas and concentration; accuracy is equal to the stated percentage.

^c Operating temperatures below -20°C may cause reduced instrument accuracy and affect display and alarm performance.

3.2 Hardware Overview

T40 II



Hardware overview

4. Operation and Instructions

The table below list that Industrial Scientific's minimum frequency recommendations for each program. These recommendations are based on site data, safe work procedures, industry best practices and regulatory standards to help ensure worker safety. Users can reduce the frequency of the following procedures as appropriate according to the safety policy of local government or company, Industrial Scientific is not responsible for developing security practices and policies.

Procedure	Recommended minimum frequency
Configuration	Before first use, when there is a change in the installed sensor type, and as otherwise needed.
Calibration	Before first use and 6month interval, depending on use and exposure to gas and contamination, such as vibration, drop or sensor replacement, multiple high concentration gas shocks,etc.
Bump test	Prior to each day's use.

4.1 Display Overview

The visual test screen depicted right shows all the indicators that can appear on the display screen. Each indicator is stationary and appears only when relevant to the task being performed.








Display screens, indicators, and abbreviations

Status indicators






- ✓ Instrument work normally.
- ! Instrument failure.
- ✂ The unit is in configuration mode.

Alarm indicators


- 🔊) The alarm icon is used in combination with other indicators to communicate a variety of conditions.
- 🔊) and ▲ High-level gas alarm.
- 🔊) and ▼ Low-level gas alarm.

 and STEL	STEL alarm.
 and TWA	TWA alarm
 and $\bar{0}r$	Positive over-range gas alarm.
 and $- \bar{0}r$	Negative over-range gas alarm.
	Low-battery alarm.


Process and time-based indicators

	The zero icon is used in combination with other indicators to communicate sensor zero information.
	The bump test icon is used in combination with other indicators to communicate bump test information.
	The calibration icon is used in combination with other indicators to communicate calibration information.
	The peak reading is the highest detected gas reading. Always clear the peak reading after calibration.
	A process is in progress. In configuration mode, indicates a time-based setting (e.g. bump test response time).

Gas name and unit-of-measure abbreviations


	Carbon Monoxide (CO)
H ₂ S	Hydrogen Sulfide (H ₂ S or H2S)
PPM	Parts per million is the unit of measure for CO and H ₂ S, PPM is the volume ratio in describing the gas concentration, 1ppm=10 ⁻⁶ equivalent to 1umol/mol.

Other abbreviations

	Vibration alarm enabled.
STEL	Short-term exposure limit, default time 15minutes. Display variations: "STEL"
TWA	Time-weighted average. default time 8h. Display variations: "TWA"

4.2 Start-up, Shutdown and Always on

Start-up


Press and hold the  for 3 seconds, the instrument starts to power on. After power on, the instrument automatically enters the visual test, firmware version, 20-second countdown and the second visual test screen, and is accompanied by warning sound and light test. After a short warm-up, it will automatically enter the real-time gas monitoring screen.



Visual test screen



Version



Countdown

Shutdown and always on



Countdown



Enter security code

Press and hold for five seconds.

After a five-second countdown:

The instrument powers off if;

- the always-on feature is disabled or
- the always-on feature is enabled and the security code is set to 000.






If this screen is activated, the shutdown process is security-code protected. To complete shutdown, the user must enter the correct three-digit code.

Value range: 000 to 999

4.3 Configuration mode

Configuration mode can only be accessed during startup, at the 20 second countdown screen, press and hold both buttons simultaneously more than 1 second to enter configuration mode, select alarm and warning related options that maximize safety within the air sampling environment.

When the unit is in configuration mode, the following apply:

- The tool icon () displays in the lower right corner of each screen.
- With successive short presses of the on-off button (), the user can scroll through the configuration loop.
- The enter button () is used to start the editing process or initiate zero.
- When editing a value, the enter button () increments the value and the on-off-mode button () saves the value.

- When editing a value, once the last value in the range is reached, the display starts again with the first value.
- When buttons (⏻) is pressed and held more than 1 seconds, the unit leaves configuration mode; it enters operation mode and the home screen is activated.
- Unless otherwise noted, when no button is pressed for 30 seconds, the unit enters operation mode and the home screen is activated.
- Any changes made in configuration mode are automatically saved to the unit and take effect immediately.

Configuration instructions

Buttons or
Screen

Button effects or Screen Description



Saves the currently displayed value or activates the next configuration mode screen.



Increase value or First press activates the value. Continued presses increment the value; hold to speed the increment pace.



Initiate zero

This screen allows the technician to complete the zero and calibration processes from configuration mode.



Low gas alarm setpoint

This screen features the status, low alarm, sensor-type, and configuration icons, with the alarm's current setpoint and unit of measure. Edit the alarm setpoint based on the following:

Value range = starts at minimum sensor measurement resolution, ends at the high gas alarm setpoint value.

Value increment = sensor measurement resolution



High gas alarm setpoint

This screen features the status, high alarm, sensor-type, and configuration icons, with the alarm's current setpoint and unit of measure. Edit the alarm setpoint based on the following:

Value range = starts at low gas alarm setpoint value, ends at the highest measurement range value of the sensor.

Value increment = sensor measurement resolution



TWA alarm setpoint

This screen features the status, alarm, sensor-type, configuration, and TWA icons, with the alarm's current setpoint and unit of measure. The alarm setpoint can be edited.

Value increment = within the sensor measurement resolution



STEL alarm setpoint

This screen features the status, alarm, sensor type, configuration, and STEL icons, with the current setpoint. The alarm setpoint can be edited.

Value increment: sensor measurement resolution



Calibration gas

This screen features the status, sensor-type, configuration, and calibration icons, with the current calibration gas setting in the main area.

This setting reflects the concentration of calibration gas that the instrument expects to read when calibrated; it should be edited to match the cylinder's gas concentration.

Value range: within the sensor measurement range

Value increment: sensor measurement resolution



Confidence indicator

This screen features the status, alarm, and configuration icons, with the selected setting value displayed in the main area. The technician can disable or enable the indicator and choose the indicator type.

When enabled, the unit will emit the selected signal every 90 seconds in operation mode.

Note: When options 1, 2, or 3 are selected, the expected battery life will be reduced.


Values:

0 = disabled

1 = audible chirp enabled

2 = LED flash enabled

3 = audible chirp and LED flash enabled



Operation-mode bump test

This screen features the status, lock, configuration, and bump test icons, with the selected setting value displayed in the main area. The technician can enable or disable this operation mode feature.

When enabled, the instrument operator can bump test the unit from operation mode.

Values:

0 = disabled , 1 = enabled




Bump test percentage

This screen features the status, configuration, and bump test icons, with the current setting value displayed in the main area. The technician can set the percentage of calibration gas to which the unit will respond.

Value range: 50% to 95%

Value increment: 1%




Bump test response-time

This screen features the status, clock, configuration, and bump test icons, with the current setting in seconds displayed in the main area. A sensor passes a bump test when it senses the specified percentage of calibration gas within the specified response time setting.

Value range: 30 to 120 seconds


Value increment: 5 second



Alarm latch

This screen features the status, alarm, lock, and configuration icons, with the current setting displayed in the main area. The technician can enable or disable this operation-mode feature.

When disabled, a unit in alarm will turn off its alarm when the gas reading is no longer at the alarm-producing concentration.

When enabled, a unit in alarm will remain in alarm until it is manually reset. The instrument operator can reset a latched alarm from gas normal reading screen by click  button.

Values:

0 = disabled , 1 = enabled



Vibration alarm

This screen features the status, alarm, "Vib", and configuration icons, with the selected setting value displayed in the main area. When enabled, the vibrating alarm will be activated when the unit is in alarm.

Values: 0 = disabled , 1 = enabled



Operation-mode zero

This screen features the status, lock, zero, and configuration icons. The technician can enable or disable this operation-mode option.

When enabled, the instrument operator can zero the unit from operation mode.

Values: 0 = disabled , 1 = enabled



Operation-mode calibration

This screen features the status, lock, configuration, and calibration icons. The technician can enable or disable this operation-mode option.

When enabled, the instrument operator can calibrate the unit from operation mode.

Values: 0 = disabled , 1 = enabled



Security code

This screen features the status, lock, and configuration icons, with the current security code displayed in the main area. The security code controls access to a unit's configuration mode and the ability to power off a unit that is configured for always-on operation.

If the security code is set at 000, entry to configuration mode is not security-code protected, and an always-on unit can be powered off without a security code. Any other value will enable the security code.

Value range: 000 to 999

Value increment: 1



Always-on

This screen features the status, lock, configuration, and the battery icon. The technician can enable or disable this feature.

When enabled, the entry of the unit's security code (if the security code is not 000) will be required to complete the shutdown process.

Values:

0 = disabled , 1 = enabled

4.4 Operation mode






In the operation mode, the worker can scroll through the operating mode loop by continuously pressing the on/off button (⏻). The zeroing, calibration, and alarm function test procedures can only be completed if these task settings are enabled for operating mode access.

TWA, STEL (if operating mode access is enabled) and peak readings can also be viewed and cleared. When the reading summary is cleared, CO and H2S value resets to zero.

In operation mode, the following apply:


- Press ⏻ to scroll through the operation-mode loop.
- Press ⬅️ to initiate a task or to clear a reading.
- A long press on ⬅️ will reset a latched alarm; it does not disable an enabled latch.
- Except where noted, when no button is pressed for 30 seconds, the home screen is activated.

Operation instructions

Buttons or Screen	Button effects or Screen Description
	Saves the currently displayed value or activates the next configuration mode screen.
	Increase value or First press activates the value. Continued presses increment the value; hold to speed the increment pace.
	Gas monitoring This screen (numeric shown) features the check mark and sensor-type icons, the current gas reading, and unit of measure. The check mark indicates the unit is operational and there are no sensor faults.
	Peak reading This screen features the check mark, peak, and sensor-type icons, and the most recent peak reading, Press ⬅️ clear the reading then the value reset to 0.
	TWA reading This screen displays when the TWA reading is enabled in operation-mode. The screen features the check mark, sensor-type, and TWA icons, and the current TWA reading. Press ⬅️ clear the reading then the value reset to 0.



STEL reading

This screen displays when the STEL reading is enabled in operation-mode. The screen features the check mark, sensor-type, and STEL icons, and the current STEL reading. Press  clear the reading then the value reset to 0 .



Initiate bump test

This screen displays when operation-mode bump testing is enabled. The screen features the check mark and bump test icons.



Initiate zero

This screen displays when operation-mode zeroing is enabled. It features the check mark and zero icons.









4.5 Bump test (or "functional test")

Bump testing is a functional test in which an instrument's installed sensors are to be briefly exposed to (or "bumped" by) calibration gases in concentrations that are greater than the sensors' low-alarm setpoints. This will cause the instrument to go into low alarm and will indicate which sensors pass or fail this basic test for response to gas.

Note: If fails the bump test, please perform a calibration. If fails the calibration, please contact Industrial Scientific after-sales service.



Bump test instructions

Screen	Screen Description
	<p>Initiate bump test</p> <p>From anywhere in the operation-mode loop, press  until the initiate bump test screen is activated.</p> <p>Press  to start the bump test process. Press  to cancel the bump test.</p>
	<p>Bump test apply gas</p> <p>Once the bump test is started, the apply-gas screen is activated; the expected type and concentration of calibration gas are displayed.</p> <p>This screen remains activated for up to 5 minutes as the unit awaits the application of calibration gas.</p>
	<p>Bump test in progress</p>
	<p>Bump test results</p> <p>If fail the bump test, need to calibrate the instrument.</p>
	<p>If pass the bump test, the home screen will be automatically activated.</p> <p><i>note: Close the cylinder and remove the calibration cup after bump test.</i></p>

4.6 Zeroing

Zeroing adjusts the sensors' "baseline" readings, which become the points of comparison for subsequent gas readings. It is a prerequisite for calibration. During zeroing, the installed sensors are to be exposed to an air sample from a zero-grade-air cylinder or ambient air that is known to be clean air. If there are gases in the air sample that are below the lowest alarm level, the instrument will read them as zero; its task is to read the air sample as clean air. The user's task is to ensure the air is clean.




Zeroing instructions


Screen

Screen Description



Initiate zero

From anywhere in the operation-mode loop, press  until the initiate-zero screen is activated.


At the initiate-zero screen, press  to start the zero process.



Zero in-progress


While the sensors are zeroed, the zero-in-progress screen is activated.


Note: Must be zeroed in fresh air.





Zero results

After the sensors are zeroed, the zero-results screen is activated, and an audible alert is emitted.

If the result is an "F" for fail, press  to reactivate the initiate-zero screen. Repeat the zero process.



If the result is a "P" for pass, press  to display the initiate-calibration screen. If calibration is not desired, press  twice or wait approximately 30 seconds to active home screen.

4.7 Calibration

Regular calibrations promote the accurate measurement of gas concentration values. During calibration, an instrument's installed sensors are to be exposed to their set concentrations of calibration gases. Based on the sensors' responses, the instrument will self-adjust to compensate for declining sensor sensitivity, which naturally occurs as the installed sensors are used or "consumed".




Calibration instructions

Screen	Screen Description
	<p>Initiate calibration press button at zero results screen to enter the initiate calibration screen, then press button to start the calibration. To cancel the calibration, press button.</p>
	<p>Calibration apply gas Once calibration is started, the apply-gas screen is activated; the expected type and concentration of calibration gas are displayed. This screen remains active for up to 5 minutes as the unit awaits the application of calibration gas. To cancel calibration, press button.</p>
	<p>Calibration in progress <i>Note: Need to open the calibration gas cylinder regulator before calibration.</i></p>
	<p>Calibration results If passes calibration, Result screens are alternately activated and displays the span reserve value. If fails calibration, the audible, visual, and vibrating alarms turn on. Results screens are alternately activated; one indicates the fail results and the other displays the span reserve value. <i>note: Close the cylinder and remove the calibration cup after calibration.</i></p>

5. Alarms and Notifications

Alarms notify the instrument operator of danger. Warnings notify of a condition that needs attention. Take seriously all alarms, warnings, and indicators, and respond to each according to company policy.

Alarms

T40 II gas monitors have alarms of two different intensities, high and low. Alarms are persistent: they turn off when the alarm-causing event is no longer detected, however, if the instrument's alarm latch is enabled, an alarm will remain on until the user presses  to turn it off.

When all alarm signals are on:

- The high alarm features the red lights, with steady sound. It is fast-paced.
- The low alarm is similar to the high alarm, but it is medium-paced.

Note: Signals (visual, audible, and vibration) vary based on instrument settings.

Alarm Level : High

Gas Present (H2S is shown.)



Gas present, over-range alarm



Gas present, negative over-range alarm



Gas present, high alarm



STEL alarm

System alarm



System alarm



Critical low-battery alarm

Alarm Level: Low

Gas Present




Gas present, low alarm



TWA alarm

Notifications

Warnings persist until the event is resolved. In some cases, an unresolved warning will become more urgent in frequency. For example, a low-battery warning that is not resolved will change to alarm status indicating a critical low-battery condition.




Low-battery warning

6. Replace the battery

When the battery has 12 to 72 hours left, T40 II display Low-battery warning notification and emit 1 beep, visual and vibration alarm every 1 minute. The users need contact Industrial Scientific or qualified person to replace the battery.

1. Using a T10 Torx screwdriver, remove all four screws from the case bottom; set aside the screws.
2. Lift the battery up from its cradle. Dispose of according to company policy.
3. Orient the replacement battery so the positive and negative ends align with the "+" and "-" cradle markings, respectively. Place the new battery into the cradle negative end first. Press down on the battery to secure it in the cradle.



Note:

- ⚠ Use RAMWAY ER14335 battery only.
- ⚠ Use Do not replace battery when an explosive atmosphere is present.
- ⚠ Work on a nonconductive surface in a well-lit area.
- ⚠ Wear grounding straps to prevent electrostatic discharge (ESD), which can cause damage to the instrument's electronics.
- ⚠ Be sure to turn off the instrument before servicing the unit or replacing the battery.

7. Warranty Policy

Industrial Scientific Corporation's portable T40 II gas monitors are guaranteed for 2years from the date of shipment, warranted to be free from defects in material and workmanship, under normal and proper use and service.

This provision is used whichever occurs first. Unless otherwise stated in the Industrial Scientific documentation accompanying the product.

8. Limitation of Liability

Industrial Scientific makes no other warranties, express or implied, including but not limited to warranties for sale or for special needs.

If products do not meet the warranty above, the user's sole remedy and Industrial Scientific's sole obligation (Industrial Scientific's sole and only action) is to replace or repair such products that do not meet the warranty, or as purchased the original price of the product.

Any other special, incidental, or consequential damages, including loss of profit or utility, whether in contract or tort, arising out of the sale, manufacture or use of any sold product referred to these terms, in no event shall Industrial Scientific be liable for claims of conduct, including strict tort liability.

Contact information

Industrial Scientific Corporation

1 Life Way Pittsburgh, PA 15205-7500 USA

Web: www.indsci.com

Phone: +1 412-788-4353 or 1-800-DETECTS (338-3287)

E-mail: info@indsci.com

Fax: +1 412-788-8353

Industrial Scientific France S.A.S.

11D Rue Willy Brandt, 62002 Arras Cedex, France

Web: www.indsci.com

Téléphone : +33 (0)1 57 32 92 61

E-mail: info@eu.indsci.com

Fax: +33 (0)1 57 32 92 67

英思科传感仪器（上海）有限公司

地址：中国上海市浦东金桥出口加工区桂桥路 290 号

邮编：201206

电话：+86 21 5899 3279

传真：+86 21 5899 3280

E-mail: iscapinfogroup@indsci.com

网址：www.indsci.com

服务热线：+86 400 820 2515

To locate a nearby distributor of our products or an Industrial Scientific service center or business office, visit us at www.indsci.com.

Rendez-vous sur notre site Web www.indsci.com, si vous voulez trouver un distributeur de nos produits près de chez vous, ou, si vous recherchez un centre de service ou un bureau Industrial Scientific.

Besuchen Sie uns unter www.indsci.com, um einen Vertriebshändler unserer Produkte oder ein Servicecenter bzw. eine Niederlassung von Industrial Scientific zu finden.

Para buscar un distribuidor local de nuestros productos o un centro de servicio u oficina comercial de Industrial Scientific, visite www.indsci.com.

如需查找就近的产品经销商或 Industrial Scientific 服务中心或业务办事处，请访问我们的网站 www.indsci.com.cn

INDUSTRIAL
SCIENTIFIC